

Calibrating Unit

New Standard Solution Report

| | | | |
|-------------------------|-----------------------|--------------|--------------|
| Equipment | Alcotest 7110 MKIII-C | Serial No.: | ARWF-0359 |
| Location: | LAWRENCE TWP POLICE | | |
| Calibration File No.: | 01362 | Calib. Date: | 12/09/2019 |
| Certification File No.: | 01363 | Cert. Date: | 12/09/2019 |
| Linearity File No.: | 01364 | Lin. Date: | 12/09/2019 |
| Solution File No.: | 01367 | Soln. Date: | 12/15/2019 |
| Sequential File No.: | 01367 | File Date: | 12/15/2019 |
| Calibrating Unit: | WET | Model No.: | CU-34 |
| Control Solution %: | 0.100% | Serial No.: | DDWJ S3-0356 |
| Solution Control Lot: | 18290 | Expires: | 09/17/2020 |
| | | Bottle No.: | 0178 |

| Function | Result | Time | Temperature | Comment(s) |
|-------------------|--------|--------|----------------|---------------------|
| | %BAC | HH:MM | Simulator (°C) | or Error(s) |
| Ambient Air Blank | 0.000% | 02:14S | | |
| Control 1 EC | 0.097% | 02:15S | 34.0°C | *** TEST PASSED *** |
| Control 1 IR | 0.098% | 02:15S | 34.0°C | *** TEST PASSED *** |
| Ambient Air Blank | 0.000% | 02:16S | | |
| Control 2 EC | 0.097% | 02:16S | 34.0°C | *** TEST PASSED *** |
| Control 2 IR | 0.099% | 02:16S | 34.0°C | *** TEST PASSED *** |
| Ambient Air Blank | 0.000% | 02:17S | | |
| Control 3 EC | 0.095% | 02:18S | 34.0°C | *** TEST PASSED *** |
| Control 3 IR | 0.099% | 02:18S | 34.0°C | *** TEST PASSED *** |
| Ambient Air Blank | 0.000% | 02:19S | | |

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

Changed By:

Last Name: HICKEY

First Name: SHANE

MI: P

Signature: _____

Shane Hickey #240

Badge No.: 240

Date: 12/15/2019