

# Calibrating Unit

## New Standard Solution Report

<b>Equipment</b>	Alcotest 7110 MKIII-C	Serial No.: ARWF-0359
Location:	LAWRENCE TWP POLICE	
Calibration File No.:	01283	Calib. Date: 09/10/2018
Certification File No.:	01284	Calib. No.: 00030
Linearity File No.:	01285	Cert. Date: 09/10/2018
Solution File No.:	01306	Cert. No.: 00026
Sequential File No.:	01306	Lin. Date: 09/10/2018
		Lin. No.: 00026
		Soln. Date: 12/30/2018
		Soln. No.: 00225
		File Date: 12/30/2018
Calibrating Unit:	WET	Model No.: CU-34
Control Solution %:	0.100%	Serial No.: DDWJ S3-0355
Solution Control Lot:	18290	Expires: 09/17/2020
		Bottle No.: 0206

Function	Result %BAC	Time HH:MM	Temperature Simulator (°C)	Comment(s) or Error(s)
Ambient Air Blank	0.000%	15:11S		
Control 1 EC	0.102%	15:12S	34.0°C	*** TEST PASSED ***
Control 1 IR	0.101%	15:12S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	15:13S		
Control 2 EC	0.100%	15:13S	34.0°C	*** TEST PASSED ***
Control 2 IR	0.100%	15:13S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	15:14S		
Control 3 EC	0.100%	15:15S	34.0°C	*** TEST PASSED ***
Control 3 IR	0.101%	15:15S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	15:15S		

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

**Changed By:**

Last Name: JANOSKI

First Name: ROBERT

MI: T

Signature: \_\_\_\_\_



Badge No.: 239

Date: 12/30/2018