

Calibrating Unit

New Standard Solution Report

Equipment	Alcotest 7110 MKIII-C	Serial No.: ARWF-0359
Location:	LAWRENCE TWP POLICE	
Calibration File No.:	01342	Calib. Date: 07/10/2019
Certification File No.:	01343	Calib. No.: 00032
Linearity File No.:	01344	Cert. Date: 07/10/2019
Solution File No.:	01360	Cert. No.: 00028
Sequential File No.:	01360	Lin. Date: 07/10/2019
		Lin. No.: 00028
		Soln. Date: 12/06/2019
		Soln. No.: 00239
		File Date: 12/06/2019
Calibrating Unit:	WET	Model No.: CU-34
Control Solution %:	0.100%	Serial No.: DDWJ S3-0355
Solution Control Lot:	18290	Expires: 09/17/2020
		Bottle No.: 0453

Function	Result	Time	Temperature	Comment(s)
	%BAC	HH:MM	Simulator (°C)	or Error(s)
Ambient Air Blank	0.000%	15:26S		
Control 1 EC	0.099%	15:27S	33.9°C	*** TEST PASSED ***
Control 1 IR	0.100%	15:27S	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	15:28S		
Control 2 EC	0.099%	15:29S	33.9°C	*** TEST PASSED ***
Control 2 IR	0.100%	15:29S	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	15:29S		
Control 3 EC	0.097%	15:30S	33.9°C	*** TEST PASSED ***
Control 3 IR	0.099%	15:30S	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	15:31S		

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

Changed By:

Last Name: WHITMORE

First Name: MICHAEL

MI: J

Signature: _____

SGT Michael Whitmore

Badge No.: 215

Date: 12/06/2019

#215