

Calibrating Unit

New Standard Solution Report

Equipment Alcotest 7110 MKIII-C Serial No.: ARWF-0359
Location: LAWRENCE TWP POLICE
Calibration File No.: 01236 Calib. Date: 10/24/2017 Calib. No.: 00028
Certification File No.: 01237 Cert. Date: 10/24/2017 Cert. No.: 00024
Linearity File No.: 01238 Lin. Date: 10/24/2017 Lin. No.: 00024
Solution File No.: 01247 Soln. Date: 12/20/2017 Soln. No.: 00210
Sequential File No.: 01247 File Date: 12/20/2017

Calibrating Unit: WET Model No.: CU-34 Serial No.: DDWJ S3-0356
Control Solution %: 0.100% Expires: 06/28/2019
Solution Control Lot: 17200 Bottle No.: 0458

Function	Result %BAC	Time HH:MM	Temperature Simulator (°C)	Comment(s) or Error(s)
Ambient Air Blank	0.000%	10:07S		
Control 1 EC	0.099%	10:07S	34.0°C	*** TEST PASSED ***
Control 1 IR	0.100%	10:07S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	10:08S		
Control 2 EC	0.099%	10:09S	34.0°C	*** TEST PASSED ***
Control 2 IR	0.101%	10:09S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	10:10S		
Control 3 EC	0.099%	10:10S	34.0°C	*** TEST PASSED ***
Control 3 IR	0.101%	10:10S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	10:11S		

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

Changed By:

Last Name: MILLER

First Name: BRUCE

MI: B

Signature: _____

Badge No.: 169

Date: 12/20/2017