

Calibrating Unit New Standard Solution Report

Equipment Alcotest 7110 MKIII-C Serial No.: ARWF-0359
Location: LAWRENCE TWP POLICE
Calibration File No.: 01312 Calib. Date: 01/30/2019 Calib. No.: 00031
Certification File No.: 01313 Cert. Date: 01/30/2019 Cert. No.: 00027
Linearity File No.: 01314 Lin. Date: 01/30/2019 Lin. No.: 00027
Solution File No.: 01326 Soln. Date: 03/04/2019 Soln. No.: 00229
Sequential File No.: 01326 File Date: 03/04/2019

Calibrating Unit: WET Model No.: CU-34 Serial No.: DDWJ S3-0356
Control Solution %: 0.100% Expires: 01/31/2020
Solution Control Lot: 18050 Bottle No.: 0599

Function	Result %BAC	Time HH:MM	Temperature Simulator (°C)	Comment(s) or Error(s)
Ambient Air Blank	0.000%	17:12S		
Control 1 EC	0.102%	17:13S	34.0°C	*** TEST PASSED ***
Control 1 IR	0.103%	17:13S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	17:14S		
Control 2 EC	0.101%	17:14S	34.0°C	*** TEST PASSED ***
Control 2 IR	0.103%	17:14S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	17:15S		
Control 3 EC	0.101%	17:16S	34.0°C	*** TEST PASSED ***
Control 3 IR	0.103%	17:16S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	17:17S		

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

Changed By:

Last Name: SIMON

First Name: STEVEN

MI: R

Signature: _____

 # 212

Badge No.: 212

Date: 03/04/2019