

Calibrating Unit

New Standard Solution Report

Equipment	Alcotest 7110 MKIII-C	Serial No.:	ARWF-0359
Location:	LAWRENCE TWP POLICE		
Calibration File No.:	01362	Calib. Date:	12/09/2019
Certification File No.:	01363	Cert. Date:	12/09/2019
Linearity File No.:	01364	Lin. Date:	12/09/2019
Solution File No.:	01373	Soln. Date:	03/06/2020
Sequential File No.:	01373	File Date:	03/06/2020
Calibrating Unit:	WET	Model No.:	CU-34
Control Solution %:	0.100%	Serial No.:	DDWJ S3-0356
Solution Control Lot:	19240	Expires:	09/09/2021
		Bottle No.:	1353

Function	Result %BAC	Time HH:MM	Temperature Simulator (°C)	Comment(s) or Error(s)
Ambient Air Blank	0.000%	11:56S		
Control 1 EC	0.100%	11:57S	34.0°C	*** TEST PASSED ***
Control 1 IR	0.098%	11:57S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	11:58S		
Control 2 EC	0.099%	11:59S	34.0°C	*** TEST PASSED ***
Control 2 IR	0.099%	11:59S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	11:59S		
Control 3 EC	0.099%	12:00S	34.0°C	*** TEST PASSED ***
Control 3 IR	0.099%	12:00S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	12:01S		

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

Changed By:

Last Name: WHITMORE

First Name: MICHAEL

MI: J

Signature: _____

SGT Michael Whitmore #215

Badge No.: 215

Date: 03/06/2020