

Alcotest 7110 Calibration Record

Equipment

Alcotest 7110 MKIII-C
Location: LAWRENCE TWP POLICE
Serial No.: ARWF-0359
Calibration File No.: 01263
Calib. Date: 04/09/2018
Calib. No.: 00029
Certification File No.: 01237
Cert. Date: 10/24/2017
Cert. No.: 00024
Linearity File No.: 01238
Lin. Date: 10/24/2017
Lin. No.: 00024
Solution File No.: 01259
Soln. Date: 03/14/2018
Soln. No.: 00213
Sequential File No.: 01263
File Date: 04/09/2018


Calibrating Unit: WET
Control Solution %: 0.100%
Solution Control Lot: 16270
Model No.: CU-34
Serial No.: DDWJ S3-0356
Expires: 10/10/2018
Bottle No.: 0355

Coordinator

Last Name: KOZIEL
First Name: BARTLOMIEJ
MI:
Signature: T. I. B. Koziel # 7041
Badge No.: 7041
Date: 04/09/2018

*Black Key Temperature Probe Serial.....# DDUJ P2-003 (BK)

*Digital NIST Temperature Measuring System Serial.....# 170297888 (BK)



Pursuant to law, and the "Chemical Breath Testing Regulations" N.J.A.C. 13:51, I am a duly appointed Breath Test Coordinator/Instructor. In my official capacity, and consistent with "Calibration Check Procedure for Alcotest 7110," as established by the Chief Forensic Scientist of the Division of State Police, I perform calibration checks on approved instruments employing infrared analysis and electrochemical analysis, when utilized in a single approved instrument as a dual system of chemical breath testing. Pursuant to, and consistent with, the current "Calibration Check Procedure for Alcotest 7110," as established by the Chief Forensic Scientist, I performed a Calibration Check on the approved instrument identified on this certificate. The results of my Calibration Check are recorded on this certificate, which consists of two parts on two pages: Part I - Control Tests; and Part II - Linearity Tests. I certify that the foregoing statements made by me are true. I am aware that if any of the foregoing statements made by me are wilfully false, I am subject to punishment.

Alcotest 7110 Calibration Certificate

Part I - Control Tests

Equipment Alcotest 7110 MKIII-C Serial No.: ARWF-0359
Location: LAWRENCE TWP POLICE
Calibration File No.: 01263 Calib. Date: 04/09/2018 Calib. No.: 00029
Certification File No.: 01264 Cert. Date: 04/09/2018 Cert. No.: 00025
Linearity File No.: 01238 Lin. Date: 10/24/2017 Lin. No.: 00024
Solution File No.: 01259 Soln. Date: 03/14/2018 Soln. No.: 00213
Sequential File No.: 01264 File Date: 04/09/2018

Calibrating Unit: WET Model No.: CU-34 Serial No.: DDWJ S3-0356
Control Solution %: 0.100% Expires: 10/10/2018
Solution Control Lot: 16270 Bottle No.: 0355

Function	Result	Time	Temperature	Comment(s)
	%BAC	HH:MM	Simulator (°C)	or Error(s)
Ambient Air Blank	0.000%	12:30D		
Control 1 EC	0.100%	12:30D	34.0°C	*** TEST PASSED ***
Control 1 IR	0.100%	12:30D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	12:31D		
Control 2 EC	0.099%	12:32D	34.0°C	*** TEST PASSED ***
Control 2 IR	0.100%	12:32D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	12:32D		
Control 3 EC	0.098%	12:33D	34.0°C	*** TEST PASSED ***
Control 3 IR	0.099%	12:33D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	12:34D		

All tests within acceptable tolerance.

Coordinator

Last Name: KOZIEL

First Name: BARTLOMIEJ

MI:

Signature: _____

T. I. B. Koziel # 7041

Badge No.: 7041

Date: 04/09/2018

Pursuant to law, and the "Chemical Breath Testing Regulations" N.J.A.C. 13:51, I am a duly appointed Breath Test Coordinator/Instructor. In my official capacity, and consistent with "Calibration Check Procedure for Alcotest 7110," as established by the Chief Forensic Scientist of the Division of State Police, I perform calibration checks on approved instruments employing infrared analysis and electrochemical analysis, when utilized in a single approved instrument as a dual system of chemical breath testing. Pursuant to, and consistent with, the current "Calibration Check Procedure for Alcotest 7110," as established by the Chief Forensic Scientist, I performed a Calibration Check on the approved instrument identified on this certificate. The results of my Calibration Check are recorded on this certificate, which consists of two parts on two pages: Part I - Control Tests; and Part II - Linearity Tests. I certify that the foregoing statements made by me are true. I am aware that if any of the foregoing statements made by me are wilfully false, I am subject to punishment.

Alcotest 7110 Calibration Certificate

Part II - Linearity Tests

Equipment Alcotest 7110 MKIII-C Serial No.: ARWF-0359
Location: LAWRENCE TWP POLICE
Calibration File No.: 01263 Calib. Date: 04/09/2018 Calib. No.: 00029
Certification File No.: 01264 Cert. Date: 04/09/2018 Cert. No.: 00025
Linearity File No.: 01265 Lin. Date: 04/09/2018 Lin. No.: 00025
Solution File No.: 01259 Soln. Date: 03/14/2018 Soln. No.: 00213
Sequential File No.: 01265 File Date: 04/09/2018

Calibrating Unit: WET Model No.: CU-34 Serial No.: DDWE S3-0196
Control Solution %: 0.040% Expires: 08/10/2019
Solution Control Lot: 17240 Bottle No.: 1286

Calibrating Unit: WET Model No.: CU-34 Serial No.: DDWE S3-0205
Control Solution %: 0.080% Expires: 08/15/2019
Solution Control Lot: 17250 Bottle No.: 0148

Calibrating Unit: WET Model No.: CU-34 Serial No.: DDRF S3-0011
Control Solution %: 0.160% Expires: 08/21/2019
Solution Control Lot: 17260 Bottle No.: 0758

Function	Result	Time	Temperature	Comment(s)
	%BAC	HH:MM	Simulator (°C)	or Error(s)
Ambient Air Blank	0.000%	12:47D		
Control 1 EC	0.042%	12:47D	34.0°C	*** TEST PASSED ***
Control 1 IR	0.040%	12:47D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	12:49D		
Control 2 EC	0.040%	12:49D	34.0°C	*** TEST PASSED ***
Control 2 IR	0.040%	12:49D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	12:51D		
Control 3 EC	0.080%	12:52D	34.0°C	*** TEST PASSED ***
Control 3 IR	0.079%	12:52D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	12:53D		
Control 4 EC	0.080%	12:54D	34.0°C	*** TEST PASSED ***
Control 4 IR	0.079%	12:54D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	12:55D		
Control 5 EC	0.160%	12:56D	34.0°C	*** TEST PASSED ***
Control 5 IR	0.159%	12:56D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	12:58D		
Control 6 EC	0.159%	12:58D	34.0°C	*** TEST PASSED ***
Control 6 IR	0.159%	12:58D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	13:00D		

All tests within acceptable tolerance.

Coordinator

Last Name: KOZIEL

First Name: BARTLOMIEJ

MI:

Signature: _____

T. I. B. Koziel # 7041

Badge No.: 7041

Date: 04/09/2018

Calibrating Unit New Standard Solution Report

Equipment	Alcotest 7110 MKIII-C	Serial No.:	ARWF-0359
Location:	LAWRENCE TWP POLICE		
Calibration File No.:	01263	Calib. Date:	04/09/2018
Certification File No.:	01264	Cert. Date:	04/09/2018
Linearity File No.:	01265	Lin. Date:	04/09/2018
Solution File No.:	01266	Soln. Date:	04/09/2018
Sequential File No.:	01266	File Date:	04/09/2018
Calibrating Unit:	WET	Model No.:	CU-34
Control Solution %:	0.100%	Serial No.:	DDWJ S3-0356
Solution Control Lot:	17200	Expires:	06/28/2019
		Bottle No.:	0574

Function	Result	Time	Temperature	Comment(s)
	%BAC	HH:MM	Simulator (°C)	or Error(s)
Ambient Air Blank	0.000%	14:06D		
Control 1 EC	0.100%	14:06D	34.0°C	*** TEST PASSED ***
Control 1 IR	0.100%	14:06D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	14:07D		
Control 2 EC	0.100%	14:08D	34.0°C	*** TEST PASSED ***
Control 2 IR	0.101%	14:08D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	14:09D		
Control 3 EC	0.099%	14:09D	34.0°C	*** TEST PASSED ***
Control 3 IR	0.100%	14:09D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	14:10D		

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

Temperature Probe Serial Number: DDWA P2-214 (BK)

Changed By:

Last Name: KOZIEL First Name: BARTLOMIEJ MI:

Signature: T. I. B. Koziel # 7041 Badge No.: 7041 Date: 04/09/2018

Dräger

Alcotest® 7110 MKIII-C

CERTIFICATE OF ACCURACY

This is to certify that the Alcotest 7110 MKIII-C has been tested for accuracy and found to be in compliance with the National Highway Traffic Safety Administration Standard for evidential breath testing devices. The Alcotest MKIII-C is compliant as a "mobile" and "nonmobile" EBT with 49 FR 48854, 49 FR 48864 and 58 FR 48705. The manufacturer recommends accuracy verification of this instrument within 12 months of the calibration date below, or sooner, according to your State Specifications.

Certification Date:

3-16-15

SERIAL NUMBER:

ARWF-0359

Draeger Safety Diagnostics, Inc.

RC



Calibration
Certificate No. 1750.01

Calibration complies with ISO/IEC
17025, ANSI/NCSL Z540-1, and 9001



Cert. No.: 4000-8483336

Traceable® Certificate of Calibration for Digital Thermometer

Manufactured for and distributed by: VWR International, LLC, Radnor Corporate Center, Bldg 1, Ste 200, 100 Matsonford Road, Radnor, PA 19087

Instrument Identification:

Model: 61220-601 S/N: 170297888 Manufacturer: Control Company

Standards/Equipment:

Description	Serial Number	Due Date	NIST Traceable Reference
Temperature Calibration Bath TC-231	A79341		
Thermistor Module	A27129	12/01/17	1000401760
Temperature Probe	5267	12/06/17	B6B30059
Temperature Calibration Bath TC-191	A42238		
Thermistor Module	A27129	12/01/17	1000401760
Temperature Probe	5202	12/19/17	B6B30058-1
Temperature Calibration Bath TC-218	A73332		
Thermistor Probe	5356	1/10/18	B7104024
Readout, Digital Thermometer	B5C344	3/12/18	B7314035
Temperature Calibration Bath TC-275	B16388		
Thermistor Probe	5357	1/06/18	B7104023
Readout, Digital Thermometer	B5C344	3/12/18	B7314035

Certificate Information:

Technician: 104 Procedure: CAL-06 Cal Date: 4/22/17 Due Date: 4/22/19
Test Conditions: 23.9°C 61.0 %RH 1012 mBar

Calibration Data: (New Instrument)

Unit(s)	Nominal	As Found	In Tol	Nominal	As Left	In Tol	Min	Max	±U	TUR
°C		N.A.		0.002	-0.001	Y	-0.048	0.052	0.010	>4:1
°C		N.A.		25.000	24.999	Y	24.950	25.050	0.010	>4:1
°C		N.A.		49.998	50.000	Y	49.948	50.048	0.010	>4:1
°C		N.A.		99.998	100.003	Y	99.948	100.048	0.010	>4:1

This Instrument was calibrated using Instruments Traceable to National Institute of Standards and Technology.

A Test Uncertainty Ratio of at least 4:1 is maintained unless otherwise stated and is calculated using the expanded measurement uncertainty. Uncertainty evaluation includes the instrument under test and is calculated in accordance with the ISO "Guide to the Expression of Uncertainty in Measurement" (GUM). The uncertainty represents an expanded uncertainty using a coverage factor k=2 to approximate a 95% confidence level. In tolerance conditions are based on test results falling within specified limits with no reduction by the uncertainty of the measurement. The results contained herein relate only to the item calibrated. This certificate shall not be reproduced except in full, without written approval of Control Company.

Nominal=Standard's Reading; As Left=Instrument's Reading; In Tol=In Tolerance; Min/Max=Acceptance Range; ±U=Expanded Measurement Uncertainty; TUR=Test Uncertainty Ratio; Accuracy= $\pm(\text{Max}-\text{Min})/2$; Min = As Left Nominal(Rounded) - Tolerance; Max = As Left Nominal(Rounded) + Tolerance; Date=MM/DD/YY

Nicol Rodriguez
Nicol Rodriguez, Quality Manager

Aaron Judice
Aaron Judice, Technical Manager

Maintaining Accuracy:


In our opinion once calibrated your Digital Thermometer should maintain its accuracy. There is no exact way to determine how long calibration will be maintained. Digital Thermometers change little, if any at all, but can be affected by aging, temperature, shock, and contamination.

Recalibration:

For factory calibration and re-certification traceable to National Institute of Standards and Technology contact Control Company.

CONTROL COMPANY 12554 Galveston RD Suite B230 Webster TX USA 77598
Phone 281 482-1714 Fax 281 482-9448 service@control3.com www.control3.com

Control Company is an ISO 17025:2005 Calibration Laboratory Accredited by (A2LA) American Association for Laboratory Accreditation, Certificate No. 1750.01.
Control Company is ISO 9001:2008 Quality Certified by (DNV) Det Norske Veritas, Certificate No. CERT-01805-2008-AQ-HOU-RvA.
International Laboratory Accreditation Cooperation (ILAC) - Multilateral Recognition Arrangement (MRA).



Dräger

Simulator

CERTIFICATE OF ACCURACY

This Certificate of Accuracy verifies that the specified unit has been examined and found to be in compliance with National Highway and Traffic Safety Administration regulations for devices used to calibrate Evidential Breath Testers.
(F.R. Vol. 59 No. 249 12/19/94 Notices)
Dräger, Inc.



- Model: ALCOTEST CU34
- Model: MARK IIA
- Other: _____

Serial Number:
DDWES3-0196

Certification Date:
9-26-17

Technician:
BC

Re-Certification Due Date:
9-26-18



Dräger

Simulator

CERTIFICATE OF ACCURACY

This Certificate of Accuracy verifies that the specified unit has been examined and found to be in compliance with National Highway and Traffic Safety Administration regulations for devices used to calibrate Evidential Breath Testers.
(F.R. Vol. 59 No. 249 12/19/94 Notices)
Dräger, Inc.

- Model: ALCOTEST CU34
- Model: MARK IIA
- Other: _____

Serial Number:
DDWES3-0205

Certification Date:
9-26-17

Technician:
BC

Re-Certification Due Date:
9-26-18



Dräger

Simulator

CERTIFICATE OF ACCURACY

This Certificate of Accuracy verifies that the specified unit has been examined and found to be in compliance with National Highway and Traffic Safety Administration regulations for devices used to calibrate Evidential Breath Testers.
(F.R. Vol. 59 No. 249 12/19/94 Notices)
Dräger, Inc.

- Model: ALCOTEST CU34
- Model: MARK IIA
- Other: _____

Serial Number:

DDRES3-0011

Certification Date:

9-26-17

Technician:

BC

Re-Certification Due Date:

9-26-18

Dräger

Alcotest 7110 Temperature Probe

CERTIFICATE OF ACCURACY

This is to certify that the Alcotest 7110 Temperature Probe has been tested for accuracy with with instrumentation that is traceable to the National Institute of Standards and Technology (NIST). The manufacturer recommends accuracy verification of the Temperature Probe within 12 months of the certification date below, or sooner, according to your state's specifications.
For accurate temperature readings, the probe value on this certificate, noted below, must be programmed into the Alcotest 7110.

Serial Number Temp Probe:

DDUSP2-003

Certification Date:

9-8-17

Next Certification Due:

9-8-18

Probe Value:

98

Dräger, Inc.

BC

Dräger

Simulator

CERTIFICATE OF ACCURACY

This Certificate of Accuracy verifies that the specified unit has been examined and found to be in compliance with National Highway and Traffic Safety Administration regulations for devices used to calibrate Evidential Breath Testers.
(F.R. Vol. 59 No. 249 12/19/94 Notices)
Draeger, Inc.

- Model: ALCOTEST CU34
- Model: MARK IIA
- Other: _____

Serial Number:

DDW153-0356

Certification Date:

9-28-17

Technician:

BC

Re-Certification Due Date:

9-28-18

Dräger

Alcotest 7110 Temperature Probe

CERTIFICATE OF ACCURACY

This is to certify that the Alcotest 7110 Temperature Probe has been tested for accuracy with instrumentation that is traceable to the National Institute of Standards and Technology (NIST). The manufacturer recommends accuracy verification of the Temperature Probe within 12 months of the certification date below, or sooner, according to your state's specifications. For accurate temperature readings, the probe value on this certificate, noted below, must be programmed into the Alcotest 7110.

Serial Number Temp Probe:

DDWAP2-214

Certification Date:

9-28-17

Next Certification Due:

9-28-18

Probe Value:

104

Draeger, Inc.

BC



State of New Jersey

OFFICE OF THE ATTORNEY GENERAL
DEPARTMENT OF LAW AND PUBLIC SAFETY

DIVISION OF STATE POLICE

POST OFFICE BOX 7068

WEST TRENTON, NJ 08628-0068

(609) 882-2000

Superintendent

COLONEL JOSEPH R. FUENTES

Attorney General

CHRISTOPHER S. POKRINO

Lt. Governor

KIM GUADAGNO

Governor

CHRIS CHRISTIE

CERTIFICATION OF ANALYSIS
0.10 PERCENT BREATH ALCOHOL SIMULATOR SOLUTION

ACCEPTANCE SPECIFICATIONS FOR BREATH ALCOHOL SIMULATOR SOLUTION: Ethyl alcohol concentration within, but not exceeding, the range of 0.1174 to 0.1246 grams per 100 milliliters of solution.

MANUFACTURER: Draeger Safety, Inc. ANALYSIS DATE: 10/19/2016

BREATH ALCOHOL SIMULATOR SOLUTION LOT NUMBER: 16270

Representative samples of the above-referenced Lot Number were tested by Gas Chromatography and found to have a mean ethyl alcohol concentration range of 0.1203 to 0.1220 grams per 100 milliliters of solution.

This lot of breath alcohol simulator solution may be utilized as a known traceable standard for the purpose of conducting periodic tests, pursuant to N.J.A.C. 13:51-4.3, of approved breath test instruments (N.J.A.C. 13:51-3.5) utilized by law enforcement agencies in this State. The manufacturer's expiration date for this lot of breath alcohol simulator solution is October 10, 2018.

As Research Scientist for the Division of State Police, I hereby certify and attest that the tests and results documented in this Certificate of Analysis were performed at the Office of Forensic Sciences of the Division of State Police on properly functioning and calibrated instruments and equipment. All procedures utilized are accurate, objective, and performed on a routine basis by personnel within the Office of Forensic Sciences, in accordance with their professional duties and responsibilities.

All M. Alaoui, Ph.D.

Research Scientist

NJSP Office of Forensic Sciences

Sworn to and subscribed before me this 20 day of October, 2016.

Notary

JOHN R LEAVER

ID # 2207138

NOTARY PUBLIC

STATE OF NEW JERSEY

My Commission Expires Dec. 14, 2017



"An Internationally Accredited Agency"
New Jersey is an Equal Opportunity Employer
Printed on Recycled Paper and Recyclable





State of New Jersey

OFFICE OF THE ATTORNEY GENERAL
DEPARTMENT OF LAW AND PUBLIC SAFETY
DIVISION OF STATE POLICE
POST OFFICE BOX 7068
WEST TRENTON, NJ 08628-0068
(609) 882-2000

CHRIS CHRISTIE
Governor

CHRISTOPHER S. PORRINO
Attorney General

KIM GUADAGNO
Lt. Governor

COLONEL JOSEPH R. FUENTES
Superintendent

CERTIFICATION OF ANALYSIS
0.04 PERCENT BREATH ALCOHOL SIMULATOR SOLUTION

ACCEPTANCE SPECIFICATIONS FOR BREATH ALCOHOL SIMULATOR SOLUTION: Ethyl alcohol concentration within, but not exceeding, the range of 0.0469 to 0.0499 grams per 100 milliliters of solution.

MANUFACTURER: Draeger Safety, Inc.

ANALYSIS DATE: 08/29/2017

BREATH ALCOHOL SIMULATOR SOLUTION LOT NUMBER: 17240

Representative samples of the above-referenced Lot Number were tested by Gas Chromatography and found to have a mean ethyl alcohol concentration range of 0.0483 to 0.0489 grams per 100 milliliters of solution.

This lot of breath alcohol simulator solution may be utilized as a known traceable standard for the purpose of conducting periodic tests, pursuant to N.J.A.C. 13:51-4.3, of approved breath test instruments (N.J.A.C. 13:51-3.5) utilized by law enforcement agencies in this State. The manufacturer's expiration date for this lot of breath alcohol simulator solution is August 10, 2019.

As Research Scientist for the Division of State Police, I hereby certify and attest that the tests and results documented in this Certificate of Analysis were performed at the Office of Forensic Sciences of the Division of State Police on properly functioning and calibrated instruments and equipment. All procedures utilized are accurate, objective, and performed on a routine basis by personnel within the Office of Forensic Sciences, in accordance with their professional duties and responsibilities.

[Handwritten signature of Ali M. Alaouie]

Ali M. Alaouie, Ph.D.
Research Scientist
NJSP Office of Forensic Sciences

Sworn to and subscribed before me this 30th day of August, 2017.

[Handwritten signature of Mary Elizabeth McLaughlin]
Notary

MARY ELIZABETH MCLAUGHLIN
ID # 2052190
NOTARY PUBLIC
STATE OF NEW JERSEY
My Commission Expires Dec. 24, 2018



"An Internationally Accredited Agency"

New Jersey Is An Equal Opportunity Employer
Printed on Recycled Paper and Recyclable





State of New Jersey

OFFICE OF THE ATTORNEY GENERAL
DEPARTMENT OF LAW AND PUBLIC SAFETY
DIVISION OF STATE POLICE
POST OFFICE BOX 7068
WEST TRENTON, NJ 08628-0068
(609) 882-2000

CHRIS CHRISTIE
Governor

KIM GUADAGNO
Lt. Governor

CHRISTOPHER S. PORRINO
Attorney General

COLONEL JOSEPH R. FUENTES
Superintendent

CERTIFICATION OF ANALYSIS
0.08 PERCENT BREATH ALCOHOL SIMULATOR SOLUTION

ACCEPTANCE SPECIFICATIONS FOR BREATH ALCOHOL SIMULATOR SOLUTION: Ethyl alcohol concentration within, but not exceeding, the range of 0.0939 to 0.0997 grams per 100 milliliters of solution.

MANUFACTURER: Draeger Safety, Inc.

ANALYSIS DATE: 09/07/2017

BREATH ALCOHOL SIMULATOR SOLUTION LOT NUMBER: 17250

Representative samples of the above-referenced Lot Number were tested by Gas Chromatography and found to have a mean ethyl alcohol concentration range of 0.0963 to 0.0973 grams per 100 milliliters of solution.

This lot of breath alcohol simulator solution may be utilized as a known traceable standard for the purpose of conducting periodic tests, pursuant to N.J.A.C. 13:51-4.3, of approved breath test instruments (N.J.A.C. 13:51-3.5) utilized by law enforcement agencies in this State. The manufacturer's expiration date for this lot of breath alcohol simulator solution is August 15, 2019.

As Research Scientist for the Division of State Police, I hereby certify and attest that the tests and results documented in this Certificate of Analysis were performed at the Office of Forensic Sciences of the Division of State Police on properly functioning and calibrated instruments and equipment. All procedures utilized are accurate, objective, and performed on a routine basis by personnel within the Office of Forensic Sciences, in accordance with their professional duties and responsibilities.

[Signature]
Ali M. Alaouie, Ph.D.
Research Scientist
NJSP Office of Forensic Sciences

Sworn to and subscribed before me this 11 day of September, 2017.

[Signature]
Notary

PETER F MURPHY IV
My Commission Expires
August 1, 2019



"An Internationally Accredited Agency"

New Jersey Is An Equal Opportunity Employer
Printed on Recycled Paper and Recyclable





State of New Jersey

OFFICE OF THE ATTORNEY GENERAL
DEPARTMENT OF LAW AND PUBLIC SAFETY
DIVISION OF STATE POLICE
POST OFFICE BOX 7068
WEST TRENTON, NJ 08628-0068
(609) 882-2000

CHRIS CHRISTIE
Governor

KIM GUADAGNO
Lt. Governor

CHRISTOPHER S. PORRINO
Attorney General

COLONEL JOSEPH R. FUENTES
Superintendent

CERTIFICATION OF ANALYSIS
0.16 PERCENT BREATH ALCOHOL SIMULATOR SOLUTION

ACCEPTANCE SPECIFICATIONS FOR BREATH ALCOHOL SIMULATOR SOLUTION: Ethyl alcohol concentration within, but not exceeding, the range of 0.1878 to 0.1994 grams per 100 milliliters of solution.

MANUFACTURER: Draeger Safety, Inc.

ANALYSIS DATE: 09/12/2017

BREATH ALCOHOL SIMULATOR SOLUTION LOT NUMBER: 17260

Representative samples of the above-referenced Lot Number were tested by Gas Chromatography and found to have a mean ethyl alcohol concentration range of 0.1937 to 0.1957 grams per 100 milliliters of solution.

This lot of breath alcohol simulator solution may be utilized as a known traceable standard for the purpose of conducting periodic tests, pursuant to N.J.A.C. 13:51-4.3, of approved breath test instruments (N.J.A.C. 13:51-3.5) utilized by law enforcement agencies in this State. The manufacturer's expiration date for this lot of breath alcohol simulator solution is August 21, 2019.


As Research Scientist for the Division of State Police, I hereby certify and attest that the tests and results documented in this Certificate of Analysis were performed at the Office of Forensic Sciences of the Division of State Police on properly functioning and calibrated instruments and equipment. All procedures utilized are accurate, objective, and performed on a routine basis by personnel within the Office of Forensic Sciences, in accordance with their professional duties and responsibilities.

[Handwritten signature]

Ali M. Alaouie, Ph.D.
Research Scientist
NJSP Office of Forensic Sciences

Sworn to and subscribed before me this 13 day of September, 2017.

[Handwritten signature]
Notary



"An Internationally Accredited Agency"
New Jersey Is An Equal Opportunity Employer
Printed on Recycled Paper and Recyclable





"An Internationally Accredited Agency"
New Jersey is an Equal Opportunity Employer
Printed on Recycled Paper and Recyclable



MARY ELIZABETH MCLAUGHLIN
NOTARY PUBLIC
ID # 2052190
STATE OF NEW JERSEY
My Commission Expires Dec. 24, 2018

MARY ELIZABETH MCLAUGHLIN

Notary
Mary Elizabeth McLaughlin
Sworn to and subscribed before me this 14th day of May, 2017.

NJSP Office of Forensic Sciences
Forensic Scientist III
Monica Trammontin

Monica Trammontin

As Forensic Scientist for the Division of State Police, I hereby certify and attest that the tests and results documented in this Certificate of Analysis were performed at the Office of Forensic Sciences of the Division of State Police on properly functioning and calibrated instruments and equipment. All procedures utilized are accurate, objective, and performed on a routine basis by personnel within the Office of Forensic Sciences, in accordance with their professional duties and responsibilities.

Representative samples of the above-referenced Lot Number were tested by Gas Chromatography and found to have a mean ethyl alcohol concentration range of 0.1208 to 0.1225 grams per 100 milliliters of solution. This lot of breath alcohol simulator solution may be utilized as a known traceable standard for the purpose of conducting periodic tests, pursuant to N.J.A.C. 13:51-4.3, of approved breath test instruments (N.J.A.C. 13:51-3.5) utilized by law enforcement agencies in this State. The manufacturer's expiration date for this lot of breath alcohol simulator solution is June 28, 2019.

BREATH ALCOHOL SIMULATOR SOLUTION LOT NUMBER: 17200

MANUFACTURER: Draeger Safety, Inc. ANALYSIS DATE: 07/13/2017

ACCEPTANCE SPECIFICATIONS FOR BREATH ALCOHOL SIMULATOR SOLUTION: Ethyl alcohol concentration within, but not exceeding, the range of 0.1174 to 0.1246 grams per 100 milliliters of solution.

**CERTIFICATION OF ANALYSIS
0.10 PERCENT BREATH ALCOHOL SIMULATOR SOLUTION**

COLONEL JOSEPH R. FUENTES
Superintendent

WEST TRENTON, NJ 08628-0068
POST OFFICE BOX 7068
(609) 882-2000

KIM GUADAGNO
Lt. Governor

CHRISTOPHER S. PORRINO
Attorney General

DEPARTMENT OF LAW AND PUBLIC SAFETY
DIVISION OF STATE POLICE
OFFICE OF THE ATTORNEY GENERAL

CHRIS CHRISTIE
Governor

State of New Jersey



DEPARTMENT OF
Traffic and Public Safety
 This is to certify that

Bartłomiej P. Koziel
 New Jersey State Police

IS QUALIFIED AND COMPETENT TO CONDUCT CRITICAL BREATH ANALYSIS PURSUANT TO CHAPTER 152 OF
 THE LAWS OF 1966 IN THE OPERATION OF THE **Aicotest 7110 MKIII-C**

A METHOD TO DETERMINE INTOXICATION
 ON IN UNDER MY HAND AT THE TIME AND PLACE DATES **28th** DAY OF **July**

THIS THIRTEEN AND **Eleven**


 STATE POLICE
 NEW JERSEY STATE POLICE


 ATTORNEY GENERAL
 STATE OF NEW JERSEY

ORIGINAL COURSE DATES

DATE	Refresher Course PLACE	INSTRUCTOR
1. 7-15-13	M.L.P.A	R. J.
2. 9-16-15	BERLEN C.P.A	C. S.
3. 9-12-17	BERLEN C.P.A	C. S.
4.		
5.		
6.		
7.		
8.		
9.		

S.P. 2006 (Rev. 03/10)

DEPARTMENT OF
Traffic and Public Safety
 This is to certify that

Bartłomiej Koziel
 Breath Test Coordinator/Instructor

IS QUALIFIED AND COMPETENT TO CONDUCT CHEMICAL BREATH ANALYSES PURSUANT TO CHAPTER 143 OF
 THE LAWS OF 1946 IN THE OPERATION OF THE **Alcotest 7110 MKIII-C**
 A METHOD TO DETERMINE INTOXICATION
 GIVEN UNDER MY HAND ATTENTION NEW JERSEY THIS 16th DAY OF December

TWO THOUSAND AND Sixteen

[Signature]
 SUPERINTENDENT
 NEW JERSEY STATE POLICE

[Signature]
 ATTORNEY GENERAL
 STATE OF NEW JERSEY

ORIGINAL COURSE DATES: _____

DATE	Refresher Course PLACE	INSTRUCTOR
1. _____	_____	_____
2. _____	_____	_____
3. _____	_____	_____
4. _____	_____	_____
5. _____	_____	_____
6. _____	_____	_____
7. _____	_____	_____
8. _____	_____	_____
9. _____	_____	_____

S.P. 2938 (Rev. 07/19)