

Calibrating Unit New Standard Solution Report

Equipment	Alcotest 7110 MKIII-C	Serial No.: ARWF-0359
Location:	LAWRENCE TWP POLICE	
Calibration File No.:	01362	Calib. Date: 12/09/2019
Certification File No.:	01363	Calib. No.: 00033
Linearity File No.:	01364	Cert. Date: 12/09/2019
Solution File No.:	01377	Cert. No.: 00029
Sequential File No.:	01377	Lin. Date: 12/09/2019
		Lin. No.: 00029
		Soln. Date: 05/01/2020
		Soln. No.: 00246
		File Date: 05/01/2020
Calibrating Unit:	WET	Model No.: CU-34
Control Solution %:	0.100%	Serial No.: DDWJ S3-0356
Solution Control Lot:	19240	Expires: 09/09/2021
		Bottle No.: 1341

Function	Result	Time	Temperature	Comment(s)
	%BAC	HH:MM	Simulator (°C)	or Error(s)
Ambient Air Blank	0.000%	11:17D		
Control 1 EC	0.099%	11:18D	34.0°C	*** TEST PASSED ***
Control 1 IR	0.099%	11:18D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	11:19D		
Control 2 EC	0.098%	11:19D	34.0°C	*** TEST PASSED ***
Control 2 IR	0.100%	11:19D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	11:20D		
Control 3 EC	0.098%	11:21D	34.0°C	*** TEST PASSED ***
Control 3 IR	0.099%	11:21D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	11:22D		

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

Changed By:

Last Name: WHITMORE

First Name: MICHAEL

MI: J

Signature: _____

SGT Michael Whitmore

Badge No.: 215

Date: 05/01/2020

#215