

Calibrating Unit New Standard Solution Report

Equipment	Alcotest 7110 MKIII-C	Serial No.: ARWF-0359
Location:	LAWRENCE TWP POLICE	
Calibration File No.:	01312	Calib. Date: 01/30/2019
Certification File No.:	01313	Calib. No.: 00031
Linearity File No.:	01314	Cert. Date: 01/30/2019
Solution File No.:	01333	Cert. No.: 00027
Sequential File No.:	01333	Lin. Date: 01/30/2019
		Lin. No.: 00027
		Soln. Date: 05/07/2019
		Soln. No.: 00231
		File Date: 05/07/2019
Calibrating Unit:	WET	Model No.: CU-34
Control Solution %:	0.100%	Serial No.: DDWJ S3-0356
Solution Control Lot:	18050	Expires: 01/31/2020
		Bottle No.: 0719

Function	Result %BAC	Time HH:MM	Temperature Simulator (°C)	Comment(s) or Error(s)
Ambient Air Blank	0.000%	03:14D		
Control 1 EC	0.103%	03:15D	34.0°C	*** TEST PASSED ***
Control 1 IR	0.103%	03:15D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	03:16D		
Control 2 EC	0.102%	03:16D	34.0°C	*** TEST PASSED ***
Control 2 IR	0.101%	03:16D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	03:17D		
Control 3 EC	0.101%	03:18D	34.0°C	*** TEST PASSED ***
Control 3 IR	0.101%	03:18D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	03:19D		

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

Changed By:

Last Name: JANOSKI

First Name: ROBERT

MI: T

Signature: _____

Badge No.: 239

Date: 05/07/2019

206