

# Calibrating Unit

## New Standard Solution Report

<b>Equipment</b>	Alcotest 7110 MKIII-C	Serial No.: ARWF-0359
Location:	LAWRENCE TWP POLICE	
Calibration File No.:	01161	Calib. Date: 12/06/2016
Certification File No.:	01162	Calib. No.: 00026
Linearity File No.:	01163	Cert. Date: 12/06/2016
Solution File No.:	01196	Cert. No.: 00022
Sequential File No.:	01196	Lin. Date: 12/06/2016
		Lin. No.: 00022
		Soln. Date: 05/14/2017
		Soln. No.: 00201
		File Date: 05/14/2017
Calibrating Unit:	WET	Model No.: CU-34
Control Solution %:	0.100%	Serial No.: DDWJ S3-0355
Solution Control Lot:	201609A	Expires: 09/07/2018
		Bottle No.: 1487

Function	Result	Time	Temperature	Comment(s)
	%BAC	HH:MM	Simulator (°C)	or Error(s)
Ambient Air Blank	0.000%	17:22D		
Control 1 EC	0.101%	17:23D	34.0°C	*** TEST PASSED ***
Control 1 IR	0.102%	17:23D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	17:23D		
Control 2 EC	0.100%	17:24D	34.0°C	*** TEST PASSED ***
Control 2 IR	0.100%	17:24D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	17:25D		
Control 3 EC	0.100%	17:25D	34.0°C	*** TEST PASSED ***
Control 3 IR	0.100%	17:25D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	17:26D		

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.


**Changed By:**

Last Name: EVERIST

First Name: THOMAS

MI: J

Signature: \_\_\_\_\_



EVERIST 193

Badge No.: 193

Date: 05/14/2017

DEPARTMENT OF  
**Motor and Public Safety**  
 This is to certify that

**Thomas J. Everist**  
**Lawrence Township**

IS QUALIFIED AND COMPETENT TO CONDUCT CHEMICAL BREATH ANALYSES PURSUANT TO CHAPTER 142 OF  
 THE LAWS OF 1966 IN THE OPERATION OF THE **ALCOTEST 7110 MK IIC**  
 AND FIND TO DETERMINE INTOXICATION.  
 WITNESSE MY HAND AT TRENTON, NEW JERSEY THIS 23rd DAY OF September

1913 THOUSAND AND **FIVE**

*[Signature]*  
 SUPERINTENDENT  
 NEW JERSEY STATE POLICE

*[Signature]*  
 ATTORNEY GENERAL  
 STATE OF NEW JERSEY

ORIGINAL COURSE DATES

	DATE	Refresher Course PLACE	INSTRUCTOR
1.	<u>12-19-07</u>	<u>OLPA</u>	<u>[Signature]</u>
2.	<u>8-6-09</u>	<u>OLPA</u>	<u>Wm Carr</u>
3.	<u>1/5/11</u>	<u>OLPA</u>	<u>Adam Stankis</u>
4.	<u>1/10/13</u>	<u>OLPA</u>	<u>Adam Stankis</u>
5.	<u>1/20/15</u>	<u>OLPA</u>	<u>Adam Stankis</u>
6.	<u>3-9-17</u>	<u>LAKENURST</u>	<u>Wm Carr</u>
7.			
8.			
9.			

S.P. 293B (Rev. 06/03)