

# Calibrating Unit

## New Standard Solution Report

<b>Equipment</b>	Alcotest 7110 MKIII-C	Serial No.:	ARWF-0359
Location:	LAWRENCE TWP POLICE		
Calibration File No.:	01263	Calib. Date:	04/09/2018
Certification File No.:	01264	Cert. Date:	04/09/2018
Linearity File No.:	01265	Lin. Date:	04/09/2018
Solution File No.:	01269	Soln. Date:	05/06/2018
Sequential File No.:	01269	File Date:	05/06/2018
Calibrating Unit:	WET	Model No.:	CU-34
Control Solution %:	0.100%	Serial No.:	DDWJ S3-0356
Solution Control Lot:	18050	Expires:	01/31/2020
		Bottle No.:	0500

Function	Result %BAC	Time HH:MM	Temperature Simulator (°C)	Comment(s) or Error(s)
Ambient Air Blank	0.000%	23:16D		
Control 1 EC	0.102%	23:17D	34.0°C	*** TEST PASSED ***
Control 1 IR	0.100%	23:17D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	23:17D		
Control 2 EC	0.100%	23:18D	34.0°C	*** TEST PASSED ***
Control 2 IR	0.100%	23:18D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	23:19D		
Control 3 EC	0.100%	23:20D	34.0°C	*** TEST PASSED ***
Control 3 IR	0.100%	23:20D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	23:20D		

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

**Changed By:**

Last Name: WHITMORE

First Name: MICHAEL

MI: J

Signature: \_\_\_\_\_

*SG Michael Whitmore #215*

Badge No.: 215

Date: 05/06/2018