

Calibrating Unit

New Standard Solution Report

Equipment	Alcotest 7110 MKIII-C	Serial No.: ARWF-0359
Location:	LAWRENCE TWP POLICE	
Calibration File No.:	01263	Calib. Date: 04/09/2018
Certification File No.:	01264	Calib. No.: 00029
Linearity File No.:	01265	Cert. Date: 04/09/2018
Solution File No.:	01270	Cert. No.: 00025
Sequential File No.:	01270	Lin. Date: 04/09/2018
		Lin. No.: 00025
		Soln. Date: 06/12/2018
		Soln. No.: 00216
		File Date: 06/12/2018
Calibrating Unit:	WET	Model No.: CU-34
Control Solution %:	0.100%	Serial No.: DDWJ S3-0356
Solution Control Lot:	18050	Expires: 01/31/2020
		Bottle No.: 0556

Function	Result	Time	Temperature	Comment(s)
	%BAC	HH:MM	Simulator (°C)	or Error(s)
Ambient Air Blank	0.000%	03:07D		
Control 1 EC	0.103%	03:08D	34.0°C	*** TEST PASSED ***
Control 1 IR	0.100%	03:08D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	03:08D		
Control 2 EC	0.101%	03:09D	34.0°C	*** TEST PASSED ***
Control 2 IR	0.100%	03:09D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	03:10D		
Control 3 EC	0.101%	03:10D	34.0°C	*** TEST PASSED ***
Control 3 IR	0.100%	03:10D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	03:11D		

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

Changed By:

Last Name: LAMONT

First Name: RICHARD

MI: N

Signature: _____

Richard Lamont #242

Badge No.: 242

Date: 06/12/2018