

Calibrating Unit

New Standard Solution Report

Equipment	Alcotest 7110 MKIII-C	Serial No.: ARWF-0359
Location:	LAWRENCE TWP POLICE	
Calibration File No.:	01378	Calib. Date: 05/15/2020
Certification File No.:	01379	Calib. No.: 00034
Linearity File No.:	01380	Cert. Date: 05/15/2020
Solution File No.:	01386	Cert. No.: 00030
Sequential File No.:	01386	Lin. Date: 05/15/2020
		Lin. No.: 00030
		Soln. Date: 06/16/2020
		Soln. No.: 00248
		File Date: 06/16/2020
Calibrating Unit:	WET	Model No.: CU-34
Control Solution %:	0.100%	Serial No.: DDWJ S3-0355
Solution Control Lot:	19240	Expires: 09/09/2021
		Bottle No.: 1364

Function	Result %BAC	Time HH:MM	Temperature Simulator (°C)	Comment(s) or Error(s)
Ambient Air Blank	0.000%	11:42D		
Control 1 EC	0.100%	11:43D	34.0°C	*** TEST PASSED ***
Control 1 IR	0.100%	11:43D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	11:44D		
Control 2 EC	0.100%	11:44D	34.0°C	*** TEST PASSED ***
Control 2 IR	0.100%	11:44D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	11:45D		
Control 3 EC	0.100%	11:46D	34.0°C	*** TEST PASSED ***
Control 3 IR	0.100%	11:46D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	11:46D		

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

Changed By:

Last Name: WHITMORE

First Name: MICHAEL

MI: J

Signature: _____

SGT Michael Whitmore #215

Badge No.: 215

Date: 06/16/2020