

Calibrating Unit New Standard Solution Report

Equipment	Alcotest 7110 MKIII-C	Serial No.:	ARWF-0359
Location:	LAWRENCE TWP POLICE	Calib. No.:	00031
Calibration File No.:	01312	Cert. Date:	01/30/2019
Certification File No.:	01313	Lin. Date:	01/30/2019
Linearity File No.:	01314	Soln. Date:	07/03/2019
Solution File No.:	01341	File Date:	07/03/2019
Sequential File No.:	01341		
Calibrating Unit:	WET	Model No.:	CU-34
Control Solution %:	0.100%	Serial No.:	DDWJ S3-0356
Solution Control Lot:	18050	Expires:	01/31/2020
		Bottle No.:	0626

Function	Result	Time	Temperature	Comment(s)
	%BAC	HH:MM	Simulator (°C)	or Error(s)
Ambient Air Blank	0.000%	14:34D		
Control 1 EC	0.102%	14:34D	34.0°C	*** TEST PASSED ***
Control 1 IR	0.102%	14:34D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	14:35D		
Control 2 EC	0.101%	14:36D	34.0°C	*** TEST PASSED ***
Control 2 IR	0.101%	14:36D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	14:37D		
Control 3 EC	0.101%	14:37D	34.0°C	*** TEST PASSED ***
Control 3 IR	0.102%	14:37D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	14:38D		

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

Changed By:

Last Name: WHITMORE

First Name: MICHAEL

MI: J

Signature: _____

Badge No.: 215

Date: 07/03/2019

#215