

# Calibrating Unit

## New Standard Solution Report

<b>Equipment</b>	Alcotest 7110 MKIII-C	Serial No.: ARWF-0359
Location:	LAWRENCE TWP POLICE	
Calibration File No.:	01378	Calib. Date: 05/15/2020
Certification File No.:	01379	Calib. No.: 00034
Linearity File No.:	01380	Cert. Date: 05/15/2020
Solution File No.:	01391	Cert. No.: 00030
Sequential File No.:	01391	Lin. Date: 05/15/2020
		Lin. No.: 00030
		Soln. Date: 07/09/2020
		Soln. No.: 00249
		File Date: 07/09/2020
Calibrating Unit:	WET	Model No.: CU-34
Control Solution %:	0.100%	Serial No.: DDWJ S3-0355
Solution Control Lot:	19240	Expires: 09/09/2021
		Bottle No.: 1370

Function	Result	Time	Temperature	Comment(s)
	%BAC	HH:MM	Simulator (°C)	or Error(s)
Ambient Air Blank	0.000%	12:39D		
Control 1 EC	0.100%	12:40D	34.0°C	*** TEST PASSED ***
Control 1 IR	0.100%	12:40D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	12:41D		
Control 2 EC	0.100%	12:41D	34.0°C	*** TEST PASSED ***
Control 2 IR	0.100%	12:41D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	12:42D		
Control 3 EC	0.100%	12:43D	34.0°C	*** TEST PASSED ***
Control 3 IR	0.100%	12:43D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	12:43D		

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

**Changed By:**

Last Name: WHITMORE

First Name: MICHAEL

MI: J

Signature:

*SGT Michael Whitmore #215*

Badge No.: 215

Date: 07/09/2020