

Calibrating Unit

New Standard Solution Report

Equipment	Alcotest 7110 MKIII-C	Serial No.:	ARWF-0359
Location:	LAWRENCE TWP POLICE		
Calibration File No.:	01378	Calib. Date:	05/15/2020
Certification File No.:	01379	Cert. Date:	05/15/2020
Linearity File No.:	01380	Lin. Date:	05/15/2020
Solution File No.:	01392	Soln. Date:	08/07/2020
Sequential File No.:	01392	File Date:	08/07/2020
Calibrating Unit:	WET	Model No.:	CU-34
Control Solution %:	0.100%	Serial No.:	DDWJ S3-0355
Solution Control Lot:	19240	Expires:	09/09/2021
		Bottle No.:	1363

Function	Result %BAC	Time HH:MM	Temperature Simulator (°C)	Comment(s) or Error(s)
Ambient Air Blank	0.000%	09:40D		
Control 1 EC	0.100%	09:41D	34.0°C	*** TEST PASSED ***
Control 1 IR	0.100%	09:41D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	09:41D		
Control 2 EC	0.100%	09:42D	34.0°C	*** TEST PASSED ***
Control 2 IR	0.101%	09:42D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	09:43D		
Control 3 EC	0.100%	09:43D	34.0°C	*** TEST PASSED ***
Control 3 IR	0.100%	09:43D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	09:44D		

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

Changed By:

Last Name: WHITMORE

First Name: MICHAEL

MI: J

Signature: _____

SGT Michael Whitmore #215

Badge No.: 215

Date: 08/07/2020