

Calibrating Unit New Standard Solution Report

Equipment	Alcotest 7110 MKIII-C	Serial No.:	ARWF-0359
Location:	LAWRENCE TWP POLICE		
Calibration File No.:	01342	Calib. Date:	07/10/2019
Certification File No.:	01343	Cert. No.:	00032
Linearity File No.:	01344	Cert. Date:	07/10/2019
Solution File No.:	01347	Lin. Date:	07/10/2019
Sequential File No.:	01347	Lin. No.:	00028
		Soln. Date:	08/17/2019
		Soln. No.:	00235
		File Date:	08/17/2019
Calibrating Unit:	WET	Model No.:	CU-34
Control Solution %:	0.100%	Serial No.:	DDWJ S3-0355
Solution Control Lot:	18050	Expires:	01/31/2020
		Bottle No.:	0730

Function	Result	Time	Temperature	Comment(s)
	%BAC	HH:MM	Simulator (°C)	or Error(s)
Ambient Air Blank	0.000%	23:45D		
Control 1 EC	0.100%	23:46D	34.0°C	*** TEST PASSED ***
Control 1 IR	0.100%	23:46D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	23:46D		
Control 2 EC	0.100%	23:47D	34.0°C	*** TEST PASSED ***
Control 2 IR	0.100%	23:47D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	23:48D		
Control 3 EC	0.100%	23:48D	34.0°C	*** TEST PASSED ***
Control 3 IR	0.100%	23:48D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	23:49D		

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

Changed By:

Last Name: WHITMORE

First Name: MICHAEL

MI: J

Signature: _____

SG Michael Whitmore
#215

Badge No.: 215

Date: 08/17/2019