

Calibrating Unit

New Standard Solution Report

Equipment	Alcotest 7110 MKIII-C	Serial No.:	ARWF-0359
Location:	LAWRENCE TWP POLICE		
Calibration File No.:	01263	Calib. Date:	04/09/2018
Certification File No.:	01264	Cert. Date:	04/09/2018
Linearity File No.:	01265	Lin. Date:	04/09/2018
Solution File No.:	01280	Soln. Date:	09/04/2018
Sequential File No.:	01280	File Date:	09/04/2018
Calibrating Unit:	WET	Model No.:	CU-34
Control Solution %:	0.100%	Serial No.:	DDWJ S3-0356
Solution Control Lot:	18050	Expires:	01/31/2020
		Bottle No.:	0506

Function	Result	Time	Temperature	Comment(s)
	%BAC	HH:MM	Simulator (°C)	or Error(s)
Ambient Air Blank	0.000%	02:03D		
Control 1 EC	0.099%	02:04D	34.0°C	*** TEST PASSED ***
Control 1 IR	0.099%	02:04D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	02:04D		
Control 2 EC	0.098%	02:05D	34.0°C	*** TEST PASSED ***
Control 2 IR	0.100%	02:05D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	02:06D		
Control 3 EC	0.099%	02:06D	34.0°C	*** TEST PASSED ***
Control 3 IR	0.099%	02:06D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	02:07D		

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

Changed By:

Last Name: LAMONT

First Name: RICHARD

MI: N

Signature: _____

Badge No.: 242

Date: 09/04/2018