

Calibrating Unit

New Standard Solution Report

Equipment	Alcotest 7110 MKIII-C	Serial No.:	ARWF-0359
Location:	LAWRENCE TWP POLICE		
Calibration File No.:	01342	Calib. Date:	07/10/2019
Certification File No.:	01343	Calib. No.:	00032
Linearity File No.:	01344	Cert. Date:	07/10/2019
Solution File No.:	01352	Cert. No.:	00028
Sequential File No.:	01352	Lin. Date:	07/10/2019
		Lin. No.:	00028
		Soln. Date:	09/23/2019
		Soln. No.:	00236
		File Date:	09/23/2019
Calibrating Unit:	WET	Model No.:	CU-34
Control Solution %:	0.100%	Serial No.:	DDWJ S3-0355
Solution Control Lot:	18290	Expires:	09/17/2020
		Bottle No.:	0587

Function	Result %BAC	Time HH:MM	Temperature Simulator (°C)	Comment(s) or Error(s)
Ambient Air Blank	0.000%	14:42D		
Control 1 EC	0.099%	14:43D	34.0°C	*** TEST PASSED ***
Control 1 IR	0.099%	14:43D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	14:44D		
Control 2 EC	0.098%	14:44D	34.0°C	*** TEST PASSED ***
Control 2 IR	0.100%	14:44D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	14:45D		
Control 3 EC	0.098%	14:46D	34.0°C	*** TEST PASSED ***
Control 3 IR	0.100%	14:46D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	14:46D		

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

Changed By:

Last Name: WHITMORE

First Name: MICHAEL

MI: J

Signature:

SGT Michael Whitmore #215

Badge No.: 215

Date: 09/23/2019