

Calibrating Unit

New Standard Solution Report

Equipment	Alcotest 7110 MKIII-C	Serial No.: ARWF-0359
Location:	LAWRENCE TWP POLICE	
Calibration File No.:	01161	Calib. Date: 12/06/2016
Certification File No.:	01162	Calib. No.: 00026
Linearity File No.:	01163	Cert. Date: 12/06/2016
Solution File No.:	01185	Cert. No.: 00022
Sequential File No.:	01185	Lin. Date: 12/06/2016
		Lin. No.: 00022
		Soln. Date: 03/20/2017
		Soln. No.: 00199
		File Date: 03/20/2017
Calibrating Unit:	WET	Model No.: CU-34
Control Solution %:	0.100%	Serial No.: DDWJ S3-0355
Solution Control Lot:	201609A	Expires: 09/07/2018
		Bottle No.: 1482

Function	Result %BAC	Time HH:MM	Temperature Simulator (°C)	Comment(s) or Error(s)
Ambient Air Blank	0.000%	15:37S		
Control 1 EC	0.101%	15:37S	34.0°C	*** TEST PASSED ***
Control 1 IR	0.101%	15:37S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	15:38S		
Control 2 EC	0.100%	15:39S	34.0°C	*** TEST PASSED ***
Control 2 IR	0.101%	15:39S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	15:40S		
Control 3 EC	0.101%	15:40S	34.0°C	*** TEST PASSED ***
Control 3 IR	0.102%	15:40S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	15:41S		

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

Changed By:

Last Name: ST CLAIR

First Name: KELLY

MI: A

Signature: _____

Kelly St Clair #201

Badge No.: 201

Date: 03/20/2017