

# Calibrating Unit New Standard Solution Report

<b>Equipment</b>	Alcotest 7110 MKIII-C	Serial No.: ARWF-0359
Location:	LAWRENCE TWP POLICE	
Calibration File No.:	01399	Calib. Date: 10/27/2020
Certification File No.:	01400	Calib. No.: 00035
Linearity File No.:	01401	Cert. Date: 10/27/2020
Solution File No.:	01414	Cert. No.: 00031
Sequential File No.:	01414	Lin. Date: 10/27/2020
		Lin. No.: 00031
		Soln. Date: 03/31/2021
		Soln. No.: 00258
		File Date: 03/31/2021
Calibrating Unit:	WET	Model No.: CU-34
Control Solution %:	0.100%	Serial No.: DDWJ S3-0356
Solution Control Lot:	19240	Expires: 09/09/2021
		Bottle No.: 1347

Function	Result %BAC	Time HH:MM	Temperature Simulator (°C)	Comment(s) or Error(s)
Ambient Air Blank	0.000%	18:26S		
Control 1 EC	0.099%	18:26S	34.0°C	*** TEST PASSED ***
Control 1 IR	0.100%	18:26S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	18:27S		
Control 2 EC	0.097%	18:28S	34.0°C	*** TEST PASSED ***
Control 2 IR	0.099%	18:28S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	18:29S		
Control 3 EC	0.098%	18:29S	34.0°C	*** TEST PASSED ***
Control 3 IR	0.099%	18:29S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	18:30S		

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

**Changed By:**

Last Name: EGGERT

First Name: MARK

MI: E

Signature:  #237

Badge No.: 237

Date: 03/31/2021