

Calibrating Unit New Standard Solution Report

Equipment	Alcotest 7110 MKIII-C	Serial No.:	ARWF-0359
Location:	LAWRENCE TWP POLICE		
Calibration File No.:	01399	Calib. Date:	10/27/2020
Certification File No.:	01400	Cert. No.:	00035
Linearity File No.:	01401	Cert. Date:	10/27/2020
Solution File No.:	01408	Lin. Date:	10/27/2020
Sequential File No.:	01408	Lin. No.:	00031
		Soln. Date:	02/01/2021
		Soln. No.:	00256
		File Date:	02/01/2021
Calibrating Unit:	WET	Model No.:	CU-34
Control Solution %:	0.100%	Serial No.:	DDWJ S3-0356
Solution Control Lot:	19240	Expires:	09/09/2021
		Bottle No.:	1323

Function	Result	Time	Temperature	Comment(s)
	%BAC	HH:MM	Simulator (°C)	or Error(s)
Ambient Air Blank	0.000%	11:09S		
Control 1 EC	0.101%	11:10S	34.0°C	*** TEST PASSED ***
Control 1 IR	0.099%	11:10S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	11:11S		
Control 2 EC	0.099%	11:12S	34.0°C	*** TEST PASSED ***
Control 2 IR	0.098%	11:12S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	11:13S		
Control 3 EC	0.099%	11:13S	34.0°C	*** TEST PASSED ***
Control 3 IR	0.099%	11:13S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	11:14S		

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

Changed By:

Last Name: WHITMORE

First Name: MICHAEL

MI: J

Signature: _____

Badge No.: 215

Date: 02/01/2021

#215