

Calibrating Unit

New Standard Solution Report

Equipment	Alcotest 7110 MKIII-C	Serial No.: ARWF-0359
Location:	LAWRENCE TWP POLICE	
Calibration File No.:	01399	Calib. Date: 10/27/2020
Certification File No.:	01400	Calib. No.: 00035
Linearity File No.:	01401	Cert. Date: 10/27/2020
Solution File No.:	01410	Cert. No.: 00031
Sequential File No.:	01410	Lin. Date: 10/27/2020
		Lin. No.: 00031
		Soln. Date: 02/22/2021
		Soln. No.: 00257
		File Date: 02/22/2021

Calibrating Unit:	WET	Model No.: CU-34	Serial No.: DDWJ S3-0356
Control Solution %:	0.100%		Expires: 09/09/2021
Solution Control Lot:	19240		Bottle No.: 0976

Function	Result	Time	Temperature	Comment(s)
	%BAC	HH:MM	Simulator (°C)	or Error(s)
Ambient Air Blank	0.000%	09:13S		
Control 1 EC	0.100%	09:13S	34.0°C	*** TEST PASSED ***
Control 1 IR	0.100%	09:13S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	09:14S		
Control 2 EC	0.099%	09:15S	34.0°C	*** TEST PASSED ***
Control 2 IR	0.101%	09:15S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	09:16S		
Control 3 EC	0.099%	09:17S	34.0°C	*** TEST PASSED ***
Control 3 IR	0.099%	09:17S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	09:18S		

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

Changed By:

Last Name: WHITMORE

First Name: MICHAEL

MI: J

Signature: _____

SGT Michael Whitmore #215

Badge No.: 215

Date: 02/22/2021